

Unit under test: HF Projector w/SG @50%

Freq, Hz	Output, Ref in mV	dB Response [120 + 20log(Response/0.0263) - 20]
5000	0.4	123.6421
6000	0.1	111.6009
7000	0.25	119.5597
8000	0.1	111.6009
9000	0.3	121.1433
10000	0.4	123.6421
12500	0.3	121.1433
15000	0.55	126.4081
17500	0.35	122.4822
20000	0.6	127.1639
25000	1.2	133.1845
30000	0.8	129.6627
35000	0.6	127.1639
40000	0.4	123.6421
45000	0.4	123.6421
50000	0.3	121.1433
60000	0.3	121.1433
70000	0.15	115.1227

Ref: B&K 8103 s/n: 1767553; calib 12/93 [-211.6dB re 1V/uPa, 26.3uV/Pa]

Amp: Crown 30-30 (+/- 0.5dB 10Hz to 20KHz)

SG: Kron Hite 1000A, Low output jack

HF Response

w/SG @ 50%

